



अर्जुनधारा नगरपालिका  
नगर कार्यपालिकाको कार्यालय  
शनिश्चरे, भापा, प्रदेश नं. १, नेपाल

प.स : ०८०।८१ (पशुपन्ध्री तथा मत्स्य विकास शाखा)  
च.न :



मिति :- २०८१/०२/२५

**विषय : दररेट पेश गर्ने सम्बन्धमा ।**


प्रस्तुत विषयमा यस कार्यालय अर्न्तगत पशुपन्ध्री तथा मत्स्य विकास शाखाको आ.ब. २०८०/८१ को स्वीकृत कार्यक्रम अनुसार आवश्यक पर्ने संलग्न स्पेसिफिकेशनका आधारमा मिल्क एनालाइजर थान (दुई) खरिद गर्नु पर्ने भएकोले यस कार्यालयको मौजुदा सूचिमा रहेका सेवासंग सम्बन्धीत इच्छुक फर्म/सप्लायरसहरुले न्यूनतम दररेट भ्याट सहित सूचना प्रकाशित भएको मितिले पांच(५)दिन भित्र (२०८१/०२/२९ गते ५:०० बजेसम्म) अर्जुनधारा नगरपालिका पशुपन्ध्री तथा मत्स्य विकास शाखा कटहर डांडामा पेश गर्नु हुन यो सूचना प्रकाशित गरिएको छ ।

तिलोचन रिमाल  
प्रमुख प्रशासकीय अधिकृत

### Technical Specification of Milk Analyser

S.N	Parameter	Measuring Range	Offered Specification
1.	Fat	From 0-15%	
2.	SNF	From 0-12%	
3.	Added Water Content	From 0-100% Repeatability $\pm 0.5\%$	
4.	Accuracy	$\pm 0.1\%$	
5.	Sample Volume/Measuring time	25 ml or less/30 Second or Less	
6.	Power Supply	210-270 Volts AC/12 Volts DC Power	
7.	Environment condition	Ambient Temperature: 0-45°C RH-30-80%	
8.	Software	For automatic information collection from analyzer and weighing scale, multi windows running, saving data.	
9.	Delivery	Arjundhara Municipality, Jhapa	
10.	Accessories Required	<p><b>Milk Weighing Machine:</b> SS weighing Platform; Capacity: 50-100kg (+/-20gm); Display Graduation: 100gm.</p> <p><b>Computer:</b> Min 16" Monitor With Min 4 GB RAM and 1 TB Hard Disk With all Necessary Serials and Ports.</p> <p><b>Voltage Stabilizer</b> With appropriate Capacity for backup of electricity fluctuations.</p> <p><b>Printer:</b> To Provide delivery slip for Volume, Fat, SNF, Added Water, Price of milk.</p>	



  
 Er. Durga Acharya  
 Mechanical Engineer  
 NEC No. 3530  
 Er. Durga Acharya  
 Mechanical Engineer  
 NEC. No. 3530

  
 पदम लाल राई  
 पशु विकास अधिकृत आठौं